



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

Gail Vasterling
Director



Jeremiah W. (Jay) Nixon
Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

SIMULATOR CALIBRATION REPORT

This is to certify that the simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance to the standards set by the Rules of Missouri Department of Health and Senior Services, 19 CSR 25-30.

SIMULATOR INFORMATION

Agency:	Missouri State Highway Patrol
Serial Number:	MP2305
Manufacturer:	Guth
Model Number:	12V500

CALIBRATION RESULTS


<u>Reference Temperature</u>	<u>Simulator Temperature</u>
34.01	34.00

This calibration was performed with
NIST-Traceable Thermometer SN: 307715

This simulator was tested by: JLC

This testing was performed: 08/05/15

This certification expires: 08/05/16

Signature of certifying DHSS Scientist: 

Name of certifying DHSS Scientist: Brian M. Lutmer



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-0400 FAX: 573-751-6010
 RFLAY MISSOURI for Hearing and Speech Impaired 1-800-735-2968 VOICE 1-800-735-2466

Gail Vasterling
 Director



Jeremiah W. (Jay) Nixon
 Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

BREATH ALCOHOL SIMULATOR TEST WORKSHEETTest Simulator Information

Agency

Email for COC

Serial Number:

Manufacturer:

Model Number:

Missouri State Highway Patrol
 Jimmy.CLEVELAND@MShp.Dps.Mo.Gov
 MP2305
 GUTH
 12V500

COPY OF LABEL PLACED
ON SIMULATOR

This simulator has been calibrated according to 19CSR 25-30.051
 SERIAL NUMBER: MP2305
 EXPIRATION DATE: 8/5/16
 DATE OF CALIBRATION: 8/5/15
 SERIAL NUMBER: 307715
 DATE OF CALIBRATION: 8/5/15
 ANALYST: JLC

NIST-Traceable Reference Thermometer Information

Serial Number:

Date of Certification:

Date of Expiration:

307715
 08/19/14
 08/19/15

Test Simulator Measurements

Readings	Reference Thermometer	Test Simulator
1	34.01	34.00
2	34.01	34.00
3	34.01	34.00
4	34.01	34.00
5	34.01	34.00

Bias (δ):

-.01

Technician performing testing:

Jimmy L. Cleveland

I hereby certify that all data submitted within this form was collected in accordance with the DHSS Procedure for the Testing of Breath Alcohol Simulators and 19 CSR 25-30.051, Breath Analyzer Calibration and Accuracy Verification Standards.

Signature:

Jimmy L. Cleveland

Date:

08/5/15

Submit completed forms for simulator certification to DHSS Breath Alcohol Program by fax at (573) 840-9139 or by email at brian.lutmer@health.mo.gov or breathalcohol@health.mo.gov.

www.health.mo.gov

Healthy Missourians for life.

The Missouri Department of Health and Senior Services will be the leader in promoting, protecting and partnering for health.

AN EQUAL OPPORTUNITY / AFFIRMATIVE ACTION EMPLOYER: Services provided on a nondiscriminatory basis.